

م/ رقم المناقصة 98/2024/85

Al-khadraa Tech الشركة ١-

1 National code (a)	Seller code (b)	Item N.O. (c)	2 Brief Description of Medical Supplies Product		3 Quantity Offered & physical unit (b)	4 Country of origin	5 Price per physical unit (Bidder may permissible Currency)				6 Total Price on DDP/Free Delivery at End-user, site.(Iraqi Dinar) 3x5(d)
			(a)				CIP price(Bidder may insert place of destination) (a)	DDP at End-user, site (b)	Incidental services as defined in schedule of Requirement (c)	DDP at End-user, site & Incidental Series (d) ={(b)+(c)}	
DIS-DE00-036	N/A	1	Quadruple plastic Blood bag sterile with CPDA solution 450ml capacity		430910	Singapore	5.46 \$	N/A	N/A	N/A	2,352,768.6 \$
DIS-DE00-037	N/A	2	Quadruple plastic Blood bag sterile with CPD solution 450ml capacity (with leukocyte filter)		475564	Singapore	13 \$	N/A	N/A	N/A	6,182,332 \$
DIS-DE25-176		3	Quadruple plastic blood bag with CPD/SAGM SOLUTION TOP AND BOTTOM 450 ML WITH IN LINE FILTER AND WITH CRYO BAG		50000	Singapore	14 \$	N/A	N/A	N/A	700,000 \$

Grand Total of Bid price: **9,235,100.6 \$**



abo dhabh drug scientific Bureau الشركة - ٢




National code (a)	Seller code (b)	Item NO. (c)	Brief Description of Medical Supplies		Quantity Offered & physical unit	Country of origin	Price per physical unit (Bidder may permissible Currency)				Total Price on DDP/Free Delivery at End-user, site.(Iraqi Dinar) 3x5(d)
			Product (a)	Unit pack size (b)			CIP price(Bidder may insert place of destination) (a)	DDP at End-user, site (b)	Incidental services as defined in schedule of Requirement (c)	DDP at End-user, site & Incidental Series (d) = {(b)+(c)}	
DIS-DE00-037	N/A	1	Quadriple plastic Blood bag sterile with CPD solution 450ml capacity (with leukocyte filter)	1	475,564	IRAN	10.90	N/A	N/A	N/A	5,183,647.50
DIS-DE00-039	N/A	2	Triple plastic Blood bag sterile with CPDA solution 450ml capacity	1	17,530	IRAN	4.40	N/A	N/A	N/A	77,132.00
DIS-DE00-036	N/A	3	Quadriple plastic Blood bag sterile with CPDA solution 450ml capacity	1	430,910	IRAN	5.40	N/A	N/A	N/A	2,326,914.00

٢٠٢٠ / ١ / ١٢

Grand Total of Bid price: [Bidders may insert permissible Currency] 7,587,693.60 \$ (In figures)

٣- شركة الحلم الطبي لتجارة الاجهزة والمستلزمات الطبية

National No :	PRODUCT	QTY (PCS)	QTY (PCS)/ CTN	PIC	Unit price (PC)	TOTAL price
DIS-DE00-036	Quadriple plastic Blood bag sterile with CPDA solution 450ml capacity	430910	45		(\$4.7)	\$2,025,277.00
DIS-DE00-037	Quadriple plastic Blood bag sterile with CPD solution 450ml capacity (with leukocyte filter)	475564	36		(\$9.81)	(\$4,665,282.84)

DIS-DE00-038	Double plastic Blood bag sterile with CPDA solution 450ml capacity	189854	80		(\$2.62)	497417.48
DIS-DE00-039	Triple plastic Blood bag sterile with CPDA solution 450ml capacity	17530	66		(\$4.01)	(\$70,295.3)
DIS-DE25-176	Quadruple plastic blood bag with CPD/SAGM SOLUTION TOP AND BOTTOM 450 ML WITH IN LINE FILTER AND WITH CRYO BAG	50000	36		(\$10.10)	(\$505,000.00)
Total in words : Seven Million seven hundred sixty three thousand two hundred seventy two US dollar and sixty two sent .					\$7,763,272.62	

٤- شركة أبراج الحياة العلمي لدعاية الادوية

n o	National Code	ITEM	Product Name	QTY	Origin-Brand	Unit Price	Total Price
1	DIS-D00-035	Single plastic blood bag disp. sterile with CPDA solution 450ml capacity	Single blood bag with sampling arm system	282411	China /Stratosphere	1.73 \$	488,571.03
2	DIS-D00-036	Quadriple plastic Blood bag sterile with CPDA solution 450ml capacity	Quadruple plastic blood bag with CPD/SAGM solution 450ml	430910	China /Stratosphere	4.15 \$	1,788,276.5
3	DIS-D00-037	Quadriple plastic Blood bag sterile with CPD solution 450ml capacity (with leukocyte filter)	Quadruple blood bag with filter for leukodepletion of RBC	475564	China /Stratosphere	8.99 \$	4,275,320.36
4	DIS-D00-038	Double blood bag with sampling arm system	Double plastic Blood bag sterile with CPDA solution 450ml capacity	189854	China /Stratosphere	2.59 \$	491,721.86
5	DIS-D00-039	Triple blood bag with sampling arm system	Triple plastic Blood bag sterile with CPDA solution 450ml capacity طلب خاص بالمركز الوطني لنقل الدم	17530	China /Stratosphere	3.97 \$	69,594.1
6	DIS-DE25-176	Quadruple blood bag Top & Bottom with filter for leukodepletion of RBC and cryo bag	Quadruple plastic blood bag with CPD/SAGM SOLUTION TOP AND BOTTOM 450 ML WITH IN LINE FILTER AND WITH CRYO BAG	50000	China /Stratosphere	9.5 \$	475,000
<b>Total</b>						<b>7,588,483.85 \$</b>	



٥- شركة الجنوة للتوريدات الطبية

SN.	KIMIADA CODE	KIMADIA	Product Name	Quantity	Price per unit S	Total Price S
1	DIS-DE00-037	Quadruple plastic Blood bag sterile with CPD solution 450ml capacity (with leukocyte filter.	<b>Blood bag with Integral Filter ( Composelect ) 4F 63ML CPD/100 ML SAG – M-RCC+ PDS-V )</b> Article Code : PQ31250	<b>475,564</b> Pcs	15.20	7,228,572.80
2	DIS-DE25-176	Quadruple plastic blood bag with CPD/SAGM SOLUTION TOP AND BOTTOM 450 ML WITH IN LINE FILTER AND WITH CRYO BAG.	<b>Blood Bag with Integral filter ( COMPOSELECT) 4F T7B 63CPD/SAG-M RCC + PDS-V ) ,( RCC T&amp;B 4F 63CPD )</b> Atrical Code :PQ32250	50000 Pcs	19.20	960,000.0
<b>Total Amount ( CIP-BAGHDAD)</b>				<b>8,188572.00S</b>		
<b>Total Amount</b>			<b>Eight Million One hundred Eighty eight thousand Five hundred Seventy two U.S DOLLARS ONLY</b>			



٦- شركة القنوان للتجارة العامة

NO.	NATIONAL CODE	ITEM	Qty	Unite Price	Total
2	DIS-DE00-036	Quadriple plastic Blood bag sterile with CPDA solution 450ml capacity	430,910	\$5.20	\$2,240,732.00
3	DIS-DE00-037	Quadriple plastic Blood bag sterile with CPD solution 450ml capacity (with leukocyte filter)	475,564	\$8.50	\$4,042,294.00
4	DIS-DE00-038	Double plastic Blood bag sterile with CPDA solution 450ml capacity	189,854	\$2.20	\$417,678.80
5	DIS-DE00-039	Triple plastic Blood bag sterile with CPDA solution 450ml capacity طلب خاص بالمركز الوطني لنقل الدم	17,530	\$3.40	\$59,602.00
المجموع النهائي فقط: ستة ملايين و سبعة و ستون ألف و ثلثمائة و ستة دولار و ثمانون سنت لا غير					\$6,760,306.80

٧- شركة AL HDBAA

NO.	NATIONAL CODE	ITEM	DESCRIPTION	UOM	QTY.	Unit Price	Total Price
1	DIS-DE00-035	Single plastic blood bag disp. sterile with CPDA solution 450ml capacity	<p>-Product name: Single blood bag with sampling arm system</p> <p>- Description</p> <p>Single bag CPDA1-450ml collection and for storage and whole blood. The system must include sampling device with vacuum holder and needle protector (for after collection)</p> <p>- Blood bag volume: 450/ ml</p> <p>-Needle and needle cap</p> <p>Size of needle "16G" with needle cap</p> <p>Once the cap opens, must be tamper-evident that the cap has been opened and not able to lock again</p> <p>Needle system should have indicator for bevel up</p> <p>- Needle protection</p> <p>The needle protection is welded. This avoids occasional opening and guarantees the product integrity</p> <p>a) Once the blood collection is completed, needle must be smoothly pulled by the operator into the</p>	PCS	282411	1.9	536580.90

2	DIS-DE00-036	Quadruple plastic Blood bag sterile with CPDA solution 450ml capacity
---	--------------	--

- Product name: Quadruple plastic blood bag with CPD/SAGM solution 450ml
- Description: Quadruple bag CPD/SAGM for preparation and storage of products (plasma, red cell and platelet and cryo. Precipitate) with blood sampling arm system adaptor for sampling with vacuum containers and with protector needle system (for needle protection after collection)
- Blood Bag volume:
Four bags 450 with CPD 63ml, 400 ml (5 days storage for platelets concentrate and gas permeable), 400ml, 400 ml(with 100 SAGM solution)
- Needle:
Size of needle "16G" with needle's cap.
Once the cap opens, must be tamper-evident that the cap has been opened and not able to lock again.
Needle system should have indicator for bevel up
- Needle protection:
The needle protection is welded. This avoids occasional opening and guarantees the product integrity
a) Once the blood collection is completed, needle must be smoothly pulled by the operator into the needle protector device (NPD) and lock. Must be signaled to personnel by an audible click or tactile indication.
b) When the collection needle is locked in the NPD, the entrance surface of NPD must extend at least 8mm beyond the tip of the needle. The collection needle should not slide out from the NPD.
c) After use, the NPD should interlock with the vacuum tube holder.
- Clamp :



				- Indications: For collection of whole blood, preparation and storage plasma , red blood cell and platelet ((5 days storage for platelets concentrate ).				
6	DIS-DE25-176	Quadruple plastic blood bag with CPD/SAGM SOLUTION TOP AND BOTTOM 450 ML WITH IN LINE FILTER AND WITH CRYO BAG	DIS-DE25-176 (Quadruple plastic blood bag with CPD/SAGM solution Top and bottom 450ml with in line filter and with (cryo bag	PCS	50000	10	500000	
			Product name	Quadruple blood bag Top & Bottom with filter for leukodepletion of RBC and cryo bag				
			Description	Quadruple bag CPD/SAGM in line filter for leukodepletion of RBC, blood sampling arm system adaptor for sampling with vacuum containers and with protector needle system (for needle protection after collection)				
			Blood Bag volume	Four bags 450 with CPD 63ml, 400 ml, 400ml, 600 ml(with 100 SAGM solution) and cryo 400ml				
			needle	Size of needle "16G" with needle's cap. Once the cap opens, must be tamper-evident that the cap has been opened and not able to lock again. Needle system should have indicator for bevel up				
			Needle protection	The needle protection is welded. This avoids occasional opening and guarantees the product integrity a) Once the blood collection is completed, needle must be smoothly pulled by the operator into the needle protector device (NPD) and lock. Must be signaled to personnel by an audible click or b) When the collection needle is locked in the NPD, the entrance surface of NPD must extend at least 8mm beyond the tip of the needle. The collection needle should not slide out from the NPD.				
5	DIS-DE00-039	Triple plastic Blood bag sterile with CPDA solution 450ml طلب خاص بالمركز الوطني لنقل الدم		-Product name: Triple blood bag with sampling arm system - Description: Triple bag CPDA1-450ml for preparation and storage of products plasma, red cell and platelet 400 ml (5 days storage for platelets concentrate and gas permeable ) with blood sampling arm system adaptor for sampling with vacuum containers and with protector needle system (for needle protection after collection) - Blood Bag volume: Three bags ( 450ml , 400 ml (5 days storage for platelets concentrate and gas permeable ) and 350 ml) - Needle: Size of needle "16G" with needle's cap. Once the cap opens, must be tamper-evident that the cap has been opened and not able to lock again. Needle system should have indicator for bevel up - Needle protection: The needle protection is welded. This avoids occasional opening and guarantees the product integrity a) Once the blood collection is completed, needle must be smoothly pulled by the operator into the needle protector device (NPD) and lock. Must be signaled to personnel by an audible click or tactile indication. b) When the collection needle is locked in the NPD, the entrance surface of NPD must extend at least 8mm beyond the tip of the needle. The collection needle should not slide out from the NPD. c) After use, the NPD should interlock with the vacuum tube holder -Clamp : Clamp is used for connecting off blood flow The sampling tube and donor tube must have	PCS	17530	4.11	72048.3

			<p>CPD 63ml / SAGM 100ml</p> <p>- Validity:</p> <p>24 month or more</p> <p>- Indications:</p> <p>Preparation of: Concentrated and filtered red blood cell CPD solution and SAGM solution is used for Storage red blood cell for 42 days and Plasma and platelets concentrate.</p>				
4	DIS-DE00-038	Double plastic Blood bag sterile with CPDA solution 450ml capacity	<p>- Product name:</p> <p>Double blood bag with sampling arm system</p> <p>- Description:</p> <p>Double bag CPDA1-450ml for preparation and storage of red blood cell and PRP products with blood sampling arm system adaptor for sampling with vacuum containers and with protector needle system (for needle protection after collection)</p> <p>- Blood Bag volume:</p> <p>Two bags ( 450ml and 350 ml)</p> <p>- Needle:</p> <p>Size of needle "16G" with needle's cap.</p> <p>Once the cap opens, must be tamper-evident that the cap has been opened and not able to lock again.</p> <p>Needle system should have indicator for bevel up</p> <p>- Needle protection:</p> <p>The needle protection is welded. This avoids occasional opening and guarantees the product integrity</p> <p>a) Once the blood collection is completed, needle must be smoothly pulled by the operator into the needle protector device (NPD) and lock. Must be signaled to personnel by an audible click or tactile indication.</p> <p>b) When the collection needle is locked in the NPD, the entrance surface of NPD must extend at least 8mm beyond the tip of the needle. The collection needle should not slide out from the NPD.</p> <p>c) After use, the NPD should interlock with the vacuum tube holder.</p> <p>- Clamp:</p> <p>Clamp is used for connecting off blood flow.</p> <p>The sampling tube and donor tube must have clamps.</p> <p>- Sampling system:</p> <p>The sampling system is completely assembled ready to accept vacuum sampler. The system is designed in order to avoid interference with the blood flow and to prevent the presence of platelets concentrate and cryo. precipitate</p>	PCS	189854	2.75	522098.5
3	DIS-DE00-037	Quadruple plastic Blood bag sterile with CPD solution 450ml capacity (with leukocyte filter)	<p>- Product name:</p> <p>Quadruple blood bag with filter for leukodepletion of RBC.</p> <p>- Description:</p> <p>Quadruple bag CPD/SAGM in line filter for leukodepletion of RBC, blood sampling arm system adaptor for sampling with vacuum containers and with protector needle system (for needle protection after collection).</p> <p>- Blood Bag volume:</p> <p>Four bags 450 with CPD 63ml, 400 ml (5 days storage for platelets concentrate and gas permeable ), 400ml , 400 ml(with 100 SAGM solution)</p> <p>- Needle:</p> <p>Size of needle "16G" with needle's cap.</p> <p>Once the cap opens, must be tamper-evident that the cap has been opened and not able to lock again.</p> <p>Needle system should have indicator for bevel up</p> <p>- Needle protection:</p> <p>The needle protection is welded. This avoids occasional opening and guarantees the product integrity</p> <p>a) Once the blood collection is completed, needle must be smoothly pulled by the operator into the needle protector device (NPD) and lock. Must be signaled to personnel by an audible click or tactile indication.</p> <p>b) When the collection needle is locked in the NPD, the entrance surface of NPD must extend at least 8mm beyond the tip of the needle. The collection needle should not slide out from the NPD.</p>	PCS	475564	10.81	5140846.84

	Each set must include secondary labels with a unique identification number (barcode). Each bag must have attached this secondary label, and additional 4 labels with the same unique identification number must be added inside the Plastic Foil. The size of the secondary labels is 4x2cm. Sampling device will include the same 4x2cm secondary label.		
	All, main or secondary labels must be tamper-proof		
Packing	Complying with ISO standard		
Filter	Filter for leukocyte removal from RBC. The mechanism of filter to removal leukocyte is ABSORPTION and blockage of leukocytes.		
Filtering matter	Biocompatible polyester		
Red cell recover	> 90%		
Plastic film	Medical grad PVC (class VI) complying with ISO specification		
Anticoagulant	CPD 63ml / SAGM 100ml		
Validity	30 month		
Indications	Preparation of: Concentrated and filtered red blood cell CPD solution and SAGM solution is used for Storage red blood cell for 42 days and Plasma and buffy and Cyro		
Total Price : Eight million six hundred forty six thousand and thirty three Dollars and four cents			8646033.04