المفاتحة المرقمة SUP 98/2025/56

١- اسم الشركة :- شركة الحلم الطبي لتجارة الاجهزة والمستلزمات الطبية

National No :	PRODUCT	QTY (PCS)	QTY (PCS)/ CTN	PIC	Unit price (PC)	TOTAL price
DIS-DE00-037	Quadruple blood bag with filter for leukodepletion of RBC	441854	50		($10.5)	($4,639,467.00)
Total in words : Four million, six hundred thirty-nine thousand, four hundred sixty-seven dollars only.					$ 4,639,467.00	

MEDICAL DREAM
FOR TRADE MEDICAL EQUIPMENT & SUPPLIES

٢- اسم الشركة :- شركة العهد للمستلزمات الطبية والمخبرية

National No :	PRODUCT	QTY (PCS)	QTY (PCS)/ CTN	Unit price (PC)	TOTAL price
DIS-DE00-037	Quadruple blood bag with filter for leukodepletion of RBC	441854	50	($10.9)	($4,816,208.60)
Total in words : Four million, eight hundred sixteen thousand, two hundred eight dollars and sixty cents					$ 4,816,208.60

٣- اسم الشركة :- شركة Dawalibi Group S.A.L

NATIONAL CODE	ITEM NO	BRIEF DESCRIPTION OF THE GOODS		QUANTITY OFFERED	COUNTRY OF ORIGIN	CIP PRICE (KIMADIA WH)	F.O.C	TOTAL PRICE
		PRODUCT	UNIT PACK SIZE					
DIS-DE00-036	1	Quadruple plastic blood bag sterile with CPDA1 solution 450 ml capacity	3 BAGS PER ALUMINIUM FOIL	10,000	CYPRUS	$5.00	-	**50,000.00**
DIS-DE00-037	2	Quadruple plastic blood bag sterile with CPD solution 450 ml capacity (with leukocyte filter)	2 BAGS PER ALUMINIUM FOIL	10,000	CYPRUS	$11.25	3%	**109,125.00**
TOTAL								**159,125.00**

GRAND TOTAL BID PRICE (IN FIGURES) $159,125.00

٤- اسم الشركة :- شركة القنوان للتجارة العامة

الخيـــــار الاول

NO.	NATIONAL CODE	ITEM	Qty	Unite Price	Total
1	DIS-DE00-036	Quadruple plastic blood bag sterile with CPDA1 solution 450ml capacity	397,200	$3.99	$1,584,828.00
2	DIS-DE00-037	Quadruple plastic blood bag sterile with CPD/ SAGM solution 450ml capacity with leukocyte filter	441,854	$8.42	$3,720,410.68
المجموع :فقط خمسة ملايين و ثلاثمائة و خمسة ألف و مائتان و ثمانية و ثلاثون دولار و ثمانية وستون سنت لا غير					$5,305,238.68

الخيـــــار الثاني

NO.	NATIONAL CODE	ITEM	Qty	Unite Price	Total
1	DIS-DE00-036	Quadruple plastic blood bag sterile with CPDA1 solution 450ml capacity	397,200	$5.10	$2,025,720.00
2	DIS-DE00-037	Quadruple plastic blood bag sterile with CPD/ SAGM solution 450ml capacity with leukocyte filter	441,854	$9.50	$4,197,613.00
المجموع : فقط ستة ملايين و مائتان و ثلاثة و عشرون ألف و ثلاثمائة و ثلاثة و ثلاثون دولار لا غير					$6,223,333.00

٥- اسم الشركة :- شركة الحدباء

Sup 98 -2025-56 - (CIP)							
NO.	NATIONAL CODE	ITEM	المواصفة الفنية	UOM	QTY.	price	total price
1	DIS-DE00-036	Quadruple plastic blood bag sterile with CPDA1 solution 450ml capacity	"- Product name: Quadruple plastic blood bag with CPDA1 solution 450ml - Description: (Name and /or Graphical symbol) Quadruple bag CPDA1 for preparation and storage of products (plasma, red cell and platelet and cryo. Precipitate) with blood sampling arm system adaptor for sampling with vacuum containers and with protector needle system (for needle protection after collection) - Blood Bag volume: Four bags 450 ml with CPDA1 63ml, 400 ml platelet gas permeable and (400ml, 400 ml) plasma bags - Needle: Size of needle "16G" with needle's cap. Once the cap opens, must be tamper-evident that the cap has been opened and not able to lock again. Needle system should have indicator for bevel up- Needle protection: The needle protection is welded. This avoids occasional opening and guarantees the product integrity a) Once the blood collection is completed, needle must be smoothly pulled by the operator into the needle protector device (NPD) and lock. Must be signaled to personnel by an audible click or tactile indication. b) When the collection needle is locked in the NPD, the entrance surface of NPD must extend at least 8mm beyond the tip of the needle . The collection needle should not slide out from the NPD. c) After use, the NPD should interlock with the vacuum tube holder 5 - Clamp: Clamp is used for connecting off blood flow. The sampling tube and donor tube must have clamps. - Sampling system: The sampling system is completely assembled ready to accept vacuum sampler. The system is designed in order to avoid interference with the blood flow and to reduce the risk of activation of the coagulation factors. Sampling pouch: Capacity of 30 - 50mL Vacuum tube holder: The barrel should be transparent and the barrel must extend at least 18mm beyond the tip of the sampling needle . The break valve should be placed after the Y connector to the donor tube (and not on the sampling pouch tube) to prevents leakage form the donor needle. - Break-off connectors: The break-off cones position is forced to let the blood components easily flow in both directions With audible click. -"Y" connectors: The ((Y)) connectors have a reduce inner volume to avoid turbulence to the blood components flow. - Labels: Complying with ISO standard 3826. Batch number and product identification code are reported also and / or in bar code system according to the ISBT standard. All, main labels must be tamper-proof 6 Name and / or Graphical symbol of the type of blood component on each bag Liable for addition stickers - Packing: Protective dual packing (individual and (Plastic or aluminum) Complying with ISO Standard 3826. - Certifications should be supplied by manufacturer's company FDA or CE MARKED -Plastic film : Medical grad PVC (class VI) complying with ISO specification 3826 (should be supplied by manufacturer's company) -Anticoagulant. CPDA1 63ml - Validity(mandatory) 24 month or more - Indications: For collection of whole blood, preparation and storage plasma, red blood cell and platelet and cryo.precipitate"&	PCS	397200	4.25	1,688,100.00

2	DIS-DE00-037	Quadruple plastic blood bag sterile with CPD/ SAGM solution 450ml capacity with leukocyte filter	"- Product name: Quadruple blood bag with filter for leukodepletion of RBC. - Description:(Name and / or Graphical symbol) Quadruple bag CPD/SAGM in line filter for leukodepletion of RBC, blood sampling arm system adaptor for sampling with vacuum containers and with protector needle system (for needle protection after collection). 7 - Blood Bags volume: Four bags 450 with CPD 63ml, 400 ml for platelets concentrate and gas permeable),(400ml plasma and 400 ml (with 100ml SAGM solution) - Needle: Size of needle "16G" with needle's cap. Once the cap opens, must be tamper-evident that the cap has been opened and not able to lock again. Needle system should have indicator for bevel up. - Needle protection: The needle protection is welded. This avoids occasional opening and guarantees the product integrity a) Once the blood collection is completed, needle must be smoothly pulled by the operator into the needle protector device (NPD) and lock. Must be signaled to personnel by an audible click or tactile indication. b) When the collection needle is locked in the NPD, the entrance surface of NPD must extend at least 8mm beyond the tip of the needle. The collection needle should not slide out from the NPD. c) After use, the NPD should interlock with the vacuum tube holder. - Clamp:Clamp is used for connecting off blood flow. The sampling tube and donor tube must have clamps. 8 -Sampling system: The sampling system is completely assembled ready to accept vacuum sample. The system is designed in order to avoid interference with the blood flow and to reduce the risk of activation of the coagulation factors. Sampling pouch: Capacity of 30 - 50mL . Vacuum tube holder: The barrel should be transparent, and the barrel must extend at least 18mm beyond the tip of the sampling needle). The break valve should be placed after the Y connector to the donor tube (and not on the sampling pouch tube) (to prevents leakage form the donor needle. - Break-off connectors: The break-off cones position is forced to let the blood components easily flow in both directions. (mandatory) With audible click -"Y" connectors: The ((Y)) connectors have a reduce inner volume to avoid turbulence to the blood components flow. - Labels(mandatory) Complying with ISO standard 3826. Batch number and product identification code are reported also and / or in bar code system according to the ISBT standard. All, main labels must be tamper-proof Name and / or Graphical symbol of the type of blood component on each bag Liable for addition stickers9 - Packing: Protective dual packing (individual and (Plastic or aluminum) Complying with ISO Standard 3826 - Filter: Filter for leukocyte removal from RBC. - Filtering matter: - Red cell recover > 90% -Certifications should be supplied by manufacturer's company FDA or CE MARKED - Plastic film: Medical grad PVC (class VI) complying with ISO specification 3826 (should be supplied by manufacturer's company) - Anticoagulant:- CPD 63ml / SAGM 100ml - Validity(mandatory) 24 month or more. - Indications: Preparation of: Concentrated and filtered red blood cell CPD solution and SAGM solution is used for Storage red blood cell for 42 days and Plasma and platelets concentrate."&	PCS	441854	10.71	4,732,256.34
total price :six million four hundred twenty thousand three hundred fifty -six and thirty four cent							6,420,356.34

٦- اسم الشركة :- شركة الجذوة للتوريدات الطبية

N.	KIMIADA CODE	KIMADIA	Product Name	Quantity	Price per unit $	Total Price $
	DIS-DE00-037	Quadruple plastic Blood bag sterile with CPD solution 450ml capacity (with leukocyte filter.	Blood bag with Integral Filter (Composelect) 4F 63ML CPD/100 ML SAG – M-RCC+ PDS-V) Article Code : PQ31250	441854 Pcs	17.5	7,732,445
Total Amount (CIP-BAGHDAD)				7,732,445		
Total Amount		Seven Million Seven hundred thirty two thousand Four hundred forty five U.S DOLLARS ONLY				